

Certificate of Analysis

Product Information:

Part No: EC1413/500mLS
Description: EC 1413 μ S/cm, 500ML

Lot code: EC143241
Expiration Date: 11-8-2017
Method of Analysis: Conductivity
Method of standardization:

*This solution is standardized using a conductivity meter that is calibrated by a NIST traceable standard. A platinized platinum cell was used for measurement .
The actual value has been determined at Hydrocal.*

Acceptance Range: 1413 μ S/cm \pm 1% @ 25° C

Actual EC value @ 25°C:

<u>Analyse</u>	<u>Result</u>
1	1410 μ S/cm

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM#</u>
Potassium Chloride	999

Certificate Date: 11-08-14

Passed by quality control:



This certificate is solely valid for the lotcode given above