

Certificate of Analysis

Product Information:

Part Number pH7.00/500mlSY
Description PH 7.00 BUFFER, COLOR-YELLOW, 500ML

Buffer lot Number pH704551
Expiration Date 5-11-2018
Method of Analysis Potentiometric
Method of standardization

*A mV instrument with a combination glass pH sensor calibrated to
N.I.S.T traceable standards was used for cross checking the solution value
The actual value has been determined at Hydrocal.*

Acceptance Range 7.00 ± 0.01 pH units @ 25° C

Actual pH value @ 25°C:

<u>Analyse</u>	<u>Result</u>
1	7,003

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM#</u>
Disodium Hydrogen Phosphate	186-II-g
Potassium Dihydrogen Phosphate	186-I-g
Potassium Hydrogen Phthalate	185i

Certificate Date: 05-11-15

Passed by quality control:



This certificate is solely valid for the lotcode given above