

Certificate of Analysis

Product Information:

Part No: EC1288/500mLS
Description: EC 12.88mS/cm, 500ML
Lot code: EC120461
Expiration Date: 25-1-2019
Method of Analysis: Conductivity
Method of standardization:

*This solution is standardized using a conductivity meter that is calibrated by a NIST traceable standard. A platinized platinum cell was used for measurement .
The actual value has been determined at Hydrocal.*

Acceptance Range: 12,88 mS/cm \pm 1% @ 25° C

Actual EC value @ 25°C:

Analyse

1

Result

12,86 mS/cm

Traceability:

PTB/N.I.S.T. Chemical

Merck KGaA, 1.01254.0500

Lot. No.

HC42399554

Certificate Date: 25-01-16

Passed by quality control:



This certificate is solely valid for the lotcode given above