

Certificate of Analysis

Product Information:

Part Number pH10.0/500mlS
Description PH 10.01 BUFFER, COLOR-BLUE, 500 ml

Buffer lot Number pH104761
Expiration Date 15-11-2019
Method of Analysis Potentiometric
Method of standardization

*A mV instrument with a combination glass pH sensor calibrated to N.I.S.T traceable standards was used for cross checking the solution value
The actual value has been determined at Hydrocal.*

Acceptance Range 10.01 ± 0.02 pH units @ 25° C

Actual pH value @ 25°C:

<u>Analyse</u>	<u>Result</u>
1	10,012

Traceability:

<u>PTB/N.I.S.T. Chemical</u>	<u>Lot. No.</u>
ZMK-CRM-pH-686-L	04/16
ZMK-CRM-pH-1001-L	15/16

Certificate Date: 15-11-16

Passed by quality control:



This certificate is solely valid for the lotcode given above