

Certificate of Analysis

Product Information:

Part Number pH9,21/500mlS
Description PH 9,21 BUFFER, 500ML

Buffer lot Number pH921671
Expiration Date 18-4-2019
Method of Analysis Potentiometric
Method of standardization

*A mV instrument with a combination glass pH sensor calibrated to N.I.S.T traceable standards was used for cross checking the solution value
The actual value has been determined at Hydrocal.*

Acceptance Range 9,21 ± 0,02 pH units @ 25° C

Actual pH value @ 25°C:

<u>Analyse</u>	<u>Result</u>
1	9,200

Traceability:

<u>PTB/N.I.S.T. Chemical</u>	<u>Lot no.</u>
ZMK-CRM-pH-918-L	`02/17
ZMK-CRM-pH-1001-L	`02/17

Certificate Date: 18-04-17

Passed by quality control:



This certificate is solely valid for the lotcode given above