



Best Service of Testing


Certificate of Compliance

Certificate Number : SHBST2013010801YEC-1

Applicant : GIORGIO BORMAC S. R. L
Via Nobel, 12 - 41012 Carpi (MO) – Italy

Manufacturer : SHANGHAI SAN-XIN INSTRUMENTATION, INC.
3F, Building No.4, No.471 Guiping Road, Shanghai

Product : 70/7 SERIES PH / CONDUCTIVITY PORTABLE METER

Trade Name : 

M/N : Model PC70, Model PH70, Model COND70, Model PC7,
Model PH7, Model COND7

Test Standard : EN 61000-6-3:2007+A1:2011
EN 61000-3-2:2006+A1:2009+A2:2009
EN 61000-3-3:2008
EN 61000-6-1:2007

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2004/108/EC. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

The certificate applies to the tested sample above mentioned only and shall not imply an assessment of the whole production. It is only valid in connection with the test report number: SHBST2013010801YER-1.

CE



**Christina
Manager**

Jan. 10, 2013

Shenzhen BST Technology Co.,Ltd.

Building No.23-24,Zhiheng Industrial Park,Guankouer Road,Nantou,NanshanDistrict,Shenzhen,Guangdong,China

Tel: 400-882-9628 E-mail:christina@bst-lab.com Complaints hotline:86-755-26747756 <http://www.bst-lab.com>